

# CALIBRATION CERTIFICATE

Issued By AcSoft Limited Calibration Laboratory

**Date Of Issue:** 04-01-2024

**Certificate No:** 1507441-3

**Calibrated By:** W. Jay

**Approved By:** A. Pullinger

## CUSTOMER

NOVA Acoustics Ltd  
Suite 13, Crown House,  
94 Armley Road,  
Armley,  
Leeds,  
LS12 2EJ

## INSTRUMENT DETAILS

**Manufacturer:** CESVA  
**Model:** CB011  
**Serial No.:** T253524  
**Description:** Acoustic Calibrator accuracy class 1 with nominal level of 94 dB, and nominal frequency of 1000 Hz

## P/O NUMBER

CONTRACT

## DATE RECEIVED

04-01-2024

## DATE CALIBRATED

04-01-2024

## ENVIRONMENTAL CONDITIONS

**Temperature:** 19.3 °C  
**Humidity:** 52.1 %rh  
**Pressure:** 99.62 kPa

## CALIBRATION RESULTS

The calibrator submitted for testing has successfully completed the Periodic tests of IEC 60942:2003 (BS EN 60942:2003) (Annex B), for class 1 sound calibrators, for the environmental conditions under which the tests were performed.

## REPORTED RESULTS

The results contained in this Certificate refer only to the measurements made at the time of test for the instrument detailed above. These results do not reflect the instrument's ability to maintain calibration.

# CALIBRATION CERTIFICATE

Issued By AcSoft Limited Calibration Laboratory

Date Of Issue: 04-01-2024

Certificate No: 1507441-3

## MEASUREMENT TRACEABILITY

The instrument under test was calibrated using the following equipment:  
Svantek SV30A Acoustic Calibrator, ACS023, Certificate No. 06397/1  
GRAS 40AG Reference Microphone, ACS009, Certificate No. 06448/2  
LAB-EL LB-706B Thermo-Barometer, ACS029, Certificate No. 1148624

## NOTES

1. The information appearing on this certificate has been compiled specifically for this instrument. This calibration certificate is produced with traceable and advanced equipment which permit comprehensive quality assurance verification of all data supplied herein.
2. The measurements in this document are traceable to GUM (Central Office of Measures), Poland
3. This calibration certificate shall not be reproduced except in full, without written permission from AcSoft Ltd.

## CALIBRATION RESULTS ARE AS FOLLOWS:

### SPL

Measured Level	Deviation From Nominal	Uncertainty	Tolerance (dB)	
dB	dB	dB	class 1	class 2
<b>94.16</b>	0.16	0.15	±0.25	±0.4

### Frequency

Measured Frequency	Deviation From Nominal	Uncertainty	Tolerance (%)	
Hz	Hz	Hz	class 1	class 2
<b>1000.01</b>	0.01	0.1	±0.7 (±7Hz)	±1.7 (±17Hz)

### THD+N

Measured Distortion	Uncertainty	Tolerance (%)	
%	%	class 1	class 2
<b>0.13</b>	0.1	2.5	3.0